



SALIVARY DHEA

ENZYME IMMUNOASSAY KIT



For Diagnostic In-Vitro Use

Item No. 1-2212, (Single) 96-Well Kit;
1-2212-5, (5-Pack) 480 Wells



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com

TABLE OF CONTENTS

Intended Use	3
Introduction	3
Test Principle	4
Safety Precautions	4
General Kit Use Advice	5
Storage	5
pH Indicator	5
Specimen Collection	6
Sample Handling and Preparation	6
Materials Supplied with Single Kit	7
Materials Needed But Not Supplied	8
Reagent Preparation	9
Procedure	10
Quality Control	11
Calculations	12
Typical Results	12
Limitations	13
Salivary DHEA Example Ranges	13
Salivary DHEA EIA Kit Performance Characteristics	14
References	17
Seller's Limited Warranty	19



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com

Intended Use

The Salimetrics™ DHEA Enzyme Immunoassay Kit is a competitive immunoassay specifically designed and validated for the *in vitro* quantitative measurement of salivary Dehydroepiandrosterone (DHEA). Salivary DHEA accurately reflects the amount of serum DHEA in circulation (1). Salimetrics has not validated this kit for use with serum or plasma samples.

Please read the complete kit insert before performing this assay. Failure to follow kit procedure and recommendations for saliva collection and sample handling may result in unreliable values.

For further information about this kit, its application, or the procedures in this insert, please contact the technical service team at Salimetrics or your local sales representative.

Introduction

DHEA is a major secretory product of the adrenal glands; it is co-released along with cortisol in response to ACTH from the pituitary gland (2). Circulating levels of DHEA decline rapidly after birth and remain low until about the age of 6-8, when they begin to rise (3). This prepubertal onset of adrenal production, known as the adrenarche, involves enzymatic changes within the adrenal gland (4). DHEA levels increase until the third decade of life, then they begin to decline again. Significant reductions occur after the age of 50 (5,6).

A major role of DHEA is to act as a circulating precursor for conversion to androgens and estrogens in tissues throughout the body, (6) and it has also been associated with immune function (7). In addition, DHEA is produced directly in the nervous system, where it functions as a neuroactive and neuroprotective factor (6,8,9).

Low levels of DHEA may occur in hypoadrenalism. High levels may occur in conditions such as 21-hydroxylase and 3 β -hydroxysteroid dehydrogenase deficiencies, virilizing adrenal adenoma and carcinoma, and some cases of female hirsutism (10,11). DHEA has also been found to be related to cases of depression, (12,13) schizophrenia, (14) bone resorption, (15,16) obesity, (17) and rheumatoid arthritis (18).

In blood, only 1-10% of DHEA is in its unbound or biologically active form. The remaining DHEA is bound to serum proteins. Unbound DHEA enters the saliva via intracellular mechanisms; the majority of DHEA in saliva is non-protein bound. Salivary DHEA levels are unaffected by salivary flow rate or salivary enzymes (1). DHEA exhibits a diurnal rhythm similar to cortisol, with highest levels in the morning after awakening, followed by a decline throughout the afternoon and evening (19).



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com

Test Principle

This is a competitive immunoassay kit. DHEA in standards and samples compete with DHEA conjugated to horseradish peroxidase for the antibody binding sites on a Microtitre Plate. After incubation, unbound components are washed away. Bound DHEA Enzyme Conjugate is measured by the reaction of the horseradish peroxidase enzyme to the substrate tetramethylbenzidine (TMB). This reaction produces a blue color. A yellow color is formed after stopping the reaction with an acidic solution. The optical density is read on a standard plate reader at 450 nm. The amount of DHEA Enzyme Conjugate detected is inversely proportional to the amount of DHEA present in the sample (20).

Safety Precautions

Read Safety Data Sheets before handling reagents.

Hazardous Ingredients

Liquid Stop Solution is caustic; use with care. We recommend the procedures listed below for all kit reagents.

Handling

Follow good laboratory practices when handling kit reagents. Laboratory coats, gloves, and safety goggles are recommended. Wipe up spills using appropriate absorbent materials while wearing protective clothing. Follow local regulations for disposal.

Emergency Exposure Measures

In case of contact, immediately wash skin or flush eyes with water for 15 minutes. Remove contaminated clothing. If inhaled, remove individual to fresh air. If individual experiences difficulty breathing call a physician.

The above information is believed to be accurate but is not all-inclusive. This information should be used only as a guide. Salimetrics will not be liable for accidents or damage resulting from misuse of product.

Safety Data Sheets are available by contacting Salimetrics at support@salimetrics.com (See www.salimetrics.com for alternative contact options).



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com

General Kit Use Advice

- This kit uses break-apart microtitre strips. You may run less than a full plate. Unused wells must be stored at 2-8°C in the foil pouch with desiccant and used in the frame provided.
- Avoid microbial contamination of opened reagents. Salimetrics recommends using opened reagents within one month. Store all reagents at 2-8°C.
- The quantity of reagent provided with a single kit is sufficient for three partial runs. The volumes of wash buffer and enzyme conjugate prepared for assays using less than a full plate should be scaled down accordingly, keeping the same dilution ratio.
- Do not mix components from different lots of kits.
- To ensure highest quality assay results, pipetting of samples and reagents must be done as quickly as possible (without interruption) across the plate. Ideally, the process should be completed within 20 minutes or less.
- When using a multichannel pipette to add reagents, always follow the same sequence when adding all reagents so that the incubation time is the same for all wells.
- When running multiple plates, or multiple sets of strips, a standard curve must be run with each individual plate and/or set of strips.
- The temperature of the laboratory may affect assays. Salimetrics' kits have been validated at 68-74°F (20-23.3°C). Higher or lower temperatures may affect OD values.
- Routine calibration of pipettes and other equipment is critical for the best possible assay performance.
- When mixing plates during assay procedures, avoid speeds that spill the contents of the wells.
- We recommend saving all reagents until data analysis has confirmed a successful run to facilitate troubleshooting if necessary.
- Prior to sample addition, please label each strip to assure plate orientation and sample order when data is acquired on plate reader.

Storage

All unopened components of this kit are stable at 2-8°C until the kit's expiration date.

pH Indicator

DHEA values from samples with a pH ≤ 4.0 or ≥ 9.0 may be inaccurate. A pH indicator in the Assay Diluent alerts the user to samples with high or low pH values. Upon addition of the Assay Diluent, acidic samples will turn yellow and alkaline samples will turn purple. Dark yellow or purple wells indicate that a pH value for that sample should be obtained using pH strips. Samples with a pH ≤ 4.0 or ≥ 9.0 should be recollected (21).



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com

Specimen Collection

Avoid sample collection within 60 minutes after eating a major meal or within 12 hours after consuming alcohol. Acidic or high sugar foods can compromise assay performance by lowering sample pH and influencing bacterial growth. To minimize these factors, rinse mouth thoroughly with water 10 minutes before sample is collected.

Collect whole saliva by unstimulated passive drool. Donors may tilt the head forward, allowing the saliva to pool on the floor of the mouth, then pass the saliva through the SalivaBio Collection Aid (SCA) into a polypropylene vial. Collection protocols/methods are available online at www.salimetrics.com or upon request.

Samples visibly contaminated with blood should be recollected. Samples may be screened for possible blood contamination (22,23) using our Blood Contamination EIA Kit (Item Nos. 1-1302/1-1302-5). Do not use dipsticks, which result in false positive values due to salivary enzymes.

It is important to record the time and date of specimen collection due to the diurnal variation in DHEA levels.

Sample Handling and Preparation

After collection, it is important to keep samples cold in order to avoid bacterial growth in the specimen. Refrigerate sample within 30 minutes, and freeze at or below -20°C within 4 hours of collection. (Samples may be stored at -20°C for up to 6 months.) For long term storage, refer to the Salimetrics Collection and Handling Advice Booklet.

Do not add sodium azide to saliva samples as a preservative, as it may cause interference in the immunoassay.

On day of assay, thaw the saliva samples completely, vortex, and centrifuge at 1500 x g for 15 minutes. Freezing saliva samples will precipitate mucins. Centrifuging removes mucins and other particulate matter which may interfere with antibody binding and affect results. Samples should be at room temperature before adding to assay plate. Pipette clear sample into appropriate wells. Re-freeze saliva samples as soon as possible after adding to the assay plate. Re-centrifuge saliva samples each time that they are thawed. Avoid multiple freeze-thaw cycles.



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com

Materials Supplied with Single Kit

	Item	Quantity/Size
1	Microtitre Plate Coated with polyclonal anti-DHEA antibodies.	1 / 96 well
2	DHEA Standard 1000 pg/mL, in a saliva-like matrix. Serially dilute before use according to Reagent Preparation. Contains: DHEA, buffer, preservative.	1 vial / 1 mL
3	DHEA Controls High, Low, in a saliva-like matrix. Ready to use. Contain: DHEA, buffer, preservative.	2 vials / 500 µL each
4	DHEA Enzyme Conjugate Concentrate. Dilute before use with Assay Diluent. (See step 5 of Procedure.) Contains: DHEA conjugated to HRP, preservative.	1 vial / 50 µL
5	Assay Diluent Contains: phosphate buffer, pH indicator, preservative.	1 bottle / 60 mL
6	Wash Buffer Concentrate (10X) Dilute before use according to Reagent Preparation. Contains: phosphate buffer, detergent, preservative.	1 bottle / 100 mL
7	TMB Substrate Solution Non-toxic, ready to use.	1 bottle / 25 mL
8	Stop Solution	1 bottle / 12.5 mL
9	Non-Specific Binding (NSB) Wells Do not contain anti-DHEA antibody. Break off and insert as blanks (optional) where needed.	1 strip
10	Adhesive Plate Covers	2



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com


Materials Needed But Not Supplied

- Precision pipette to deliver 12 μL , 50 μL , 100 μL and 150 μL
- Precision multichannel pipette to deliver 50 μL , 150 μL , and 200 μL
- Vortex
- Plate rotator with 0.08-0.17 inch orbit capable of 500 rpm
- Plate reader with 450 nm and 490 to 492 nm reference filters
- Computer software for data reduction
- Deionized water
- Reagent reservoirs
- One disposable polypropylene tube to hold at least 20 mL
- Five small disposable polypropylene tubes for dilution of standard
- Pipette tips
- Serological pipette to deliver up to 18 mL
- Centrifuge capable of 1500 x g



Reagent Preparation

- Bring all reagents to room temperature and mix before use. A minimum of 1.5 hours is recommended for the 18 mL of Assay Diluent used in Step 5 (conjugate dilution) to come to room temperature.
- Bring Microtitre Plate to room temperature before use. ***It is important to keep the foil pouch with the plate strips closed until warmed to room temperature, as humidity may have an effect on the coated wells.***
- Prepare 1X wash buffer by diluting Wash Buffer Concentrate (10X) 10-fold with room-temperature deionized water (100 mL of Wash Buffer Concentrate (10X) to 900 mL of deionized H₂O). ***Dilute only enough for current day's use and discard any leftover reagent.*** (If precipitate has formed in the concentrated wash buffer, it may be heated to 40°C for 15 minutes. Cool to room temperature before use in assay.)
- Prepare serial dilutions of the DHEA Standard as follows:
 - Label five polypropylene microcentrifuge tubes or other small tubes 2 through 6.
 - Pipette 150 µL of Assay Diluent into tubes 2 through 6.
 - Serially dilute the standard 2.5X by adding 100 µL of the 1000 pg/mL standard (tube 1) to tube 2. Mix well.
 - After changing pipette tips, remove 100 µL from tube 2 to tube 3. Mix well.
 - Continue for tubes 4, 5, and 6.
 - The final concentrations of standards for tubes 1 through 6 are, respectively, 1000 pg/mL, 400 pg/mL, 160 pg/mL, 64 pg/mL, 25.6 pg/mL, and 10.2 pg/mL. Standard concentrations in nmol/L are 3.47, 1.39, 0.55, 0.22, 0.09, and 0.03, respectively.



Procedure

Step 1: Read and prepare reagents according to the Reagent Preparation section before beginning assay. Determine your plate layout. Here is a suggested layout. (Standards, controls, and saliva samples should be assayed in duplicate.)

	1	2	3	4	5	6	7	8	9	10	11	12
A	1000 Std	1000 Std	Ctrl-H	Ctrl-H								
B	400 Std	400 Std	Ctrl-L	Ctrl-L								
C	160 Std	160 Std	SMP-1	SMP-1								
D	64 Std	64 Std	SMP-2	SMP-2								
E	25.6 Std	25.6 Std	SMP-3	SMP-3								
F	10.2 Std	10.2 Std	SMP-4	SMP-4								
G	Zero	Zero	SMP-5	SMP-5								
H	NSB*	NSB*	SMP-6	SMP-6								

*NSB = Non-specific binding wells. These may serve as blanks. Use is optional.

Step 2: Keep the desired number of strips in the strip holder and place the remaining strips back in the foil pouch. If you choose to place non-specific binding wells in H-1, 2, remove strips 1 and 2 from the strip holder and break off the bottom wells. Place the strips back into the strip holder leaving H-1, 2 blank. Break off 2 NSB wells from the strip of NSB wells included in the foil pouch. Place in H-1, 2. Alternatively, NSBs may be placed wherever you choose on the plate. Reseal the foil pouch with unused wells and desiccant. Store at 2-8°C.

Cautions: *1. Extra NSB wells should not be used for determination of standards, controls, or unknowns.*
2. Do not insert wells from one plate into a different plate.

Step 3: Pipette 18 mL of Assay Diluent into the disposable tube. (Scale down proportionally if using less than the entire plate.) Set aside for Step 5.

Step 4:

- Pipette 50 µL of standards, controls, and saliva samples into appropriate wells.
- Pipette 50 µL of Assay Diluent into 2 wells to serve as the zero.
- Pipette 50 µL of Assay Diluent into each NSB well.



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com

Step 5: Dilute the Enzyme Conjugate 1:1500 by adding 12 μL of the conjugate to the 18 mL tube of Assay Diluent. (Scale down proportionally if not using the entire plate.) Conjugate tube may be centrifuged for a few minutes to bring the liquid down to the tube bottom. Immediately mix the diluted conjugate solution and add 150 μL to each well using a multichannel pipette.

Step 6: Place adhesive cover provided over plate. Mix plate on a plate rotator for 5 minutes at 500 rpm and incubate at room temperature for a total of 3 hours.

Step 7: Wash the plate 4 times with 1X wash buffer. A plate washer is recommended. However, washing may be done by gently squirting wash buffer into each well with a squirt bottle, or by pipetting 300 μL of wash buffer into each well and then discarding the liquid over a sink. After each wash, the plate should be thoroughly blotted on paper towels before turning upright. If using a plate washer, blotting is still recommended after the last wash.

Step 8: Add 200 μL of TMB Substrate Solution to each well with a multichannel pipette.

Step 9: Mix on a plate rotator for 5 minutes at 500 rpm and incubate the plate in the dark (covered) at room temperature for an additional 25 minutes.

Step 10: Add 50 μL of Stop Solution with a multichannel pipette.

Step 11:

- Mix on a plate rotator for 3 minutes at 500 rpm. If green color remains, continue mixing until green color turns to yellow. Be sure all wells have turned yellow.

Caution: Spillage may occur if mixing speed exceeds 600 rpm.

- Wipe off bottom of plate with a water-moistened, lint-free cloth and wipe dry.
- Read in a plate reader at 450 nm. Read plate within 10 minutes of adding Stop Solution. (For best results, a secondary filter correction at 490 to 492 nm is recommended.)

Quality Control

The Salimetrics' High and Low DHEA Controls should be run with each assay. The control ranges established at Salimetrics are to be used as a guide. Each laboratory should establish its own range. Variations between laboratories may be caused by differences in techniques and instrumentation.



Calculations

1. Compute the average optical density (OD) for all duplicate wells.
2. Subtract the average OD for the NSB wells (if used) from the OD of the zero, standards, controls, and saliva samples.
3. Calculate the percent bound (B/Bo) for each standard, control, and saliva sample by dividing the OD of each well (B) by the average OD for the zero (Bo). (The zero is not a point on the standard curve.)
4. Determine the concentrations of the controls and saliva samples by interpolation using data reduction software. We recommend using a 4-parameter non-linear regression curve fit.
5. Samples with DHEA values greater than 1000 pg/mL should be diluted with Assay Diluent and rerun for accurate results. If a dilution of the sample is used, multiply the assay results by the dilution factor.

A new Standard Curve must be run with each full or partial plate.

Typical Results


The results shown below are for illustration only and should not be used to calculate results from another assay.

Well	Standard	Average OD	B	B/Bo	DHEA (pg/mL)
A1,A2	S1	0.344	0.295	0.234	1000
B1,B2	S2	0.513	0.464	0.368	400
C1,C2	S3	0.741	0.692	0.548	160
D1,D2	S4	0.958	0.909	0.720	64
E1,E2	S5	1.145	1.096	0.868	25.6
F1,F2	S6	1.230	1.181	0.936	10.2
G1,G2	Bo	1.311	1.262	NA	NA
H1,H2	NSB	0.049	NA	NA	NA



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com

Example: DHEA 4-Parameter Curve Fit



Limitations

- Linearity is established from 500 to 20 pg/mL.
- Samples with DHEA values greater than 1000 pg/mL should be diluted with Assay Diluent and rerun for accurate results. To obtain the final DHEA concentration, multiply the concentration of the diluted sample by the dilution factor.
- A pH value should be obtained on samples that appear yellow or purple after the diluted conjugate solution is added and the plate is mixed (Step 6). Samples with pH values ≤ 4.0 or ≥ 9.0 should be recollected.
- See "Specimen Collection" recommendations to ensure proper collection of saliva specimens and to avoid interfering substances.
- Samples collected with sodium azide are unsuitable for this assay.
- Any quantitative results indicating abnormal DHEA levels should be followed by additional testing and evaluation.

Salivary DHEA Example Ranges*

Group	N	Range (pg/mL)	Mean (pg/mL)	Standard Deviation (pg/mL)
Adult Females	19	22.4 - 308.8	165.6	71.6
Adult Males	20	15.9 - 291.1	153.5	68.8

*To be used as a guide only. Each laboratory should establish its own range.



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com

Salivary DHEA EIA Kit Performance Characteristics

Precision

The intra-assay precision was determined from the mean of 12 replicates each.

Saliva Sample	N	Mean (pg/mL)	Standard Deviation (pg/mL)	Coefficient of Variation (%)
H	12	618.61	32.79	5.3
L	12	44.59	2.58	5.8

The inter-assay precision was determined from the mean of average duplicates for 12 separate runs.

Saliva Sample	N	Mean (pg/mL)	Standard Deviation (pg/mL)	Coefficient of Variation (%)
H	12	579.47	45.91	7.9
L	12	34.83	2.97	8.5

Recovery

Six saliva samples containing different levels of an endogenous DHEA were spiked with known quantities of DHEA and assayed.

Saliva Sample	Endogenous (pg/mL)	Added (pg/mL)	Expected (pg/mL)	Observed (pg/mL)	Recovery (%)
1	84.45	50	134.45	136.21	101.3
2	66.84	400	466.84	511.34	109.5
3	317.92	50	367.92	334.18	90.8
4	317.92	500	817.92	919.40	112.4
5	41.84	16	57.84	64.49	111.5
6	185.04	16	201.04	181.30	90.2



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com

Sensitivity

Analytical Sensitivity

The lower limit of sensitivity was determined by interpolating the mean optical density minus 2 SDs of 10 sets of duplicates at the 0 pg/mL level. The minimal concentration of DHEA that can be distinguished from 0 is 5 pg/mL.

Functional Sensitivity

The functional sensitivity was determined by assaying 60 samples at a concentration level resulting in a CV of approximately 20%. The functional sensitivity of the salivary DHEA ELISA is 8.32 pg/mL.

Correlation with Serum

The correlation between serum and saliva DHEA was determined by assaying 39 matched samples using the Diagnostic Systems Laboratories serum DHEA radioimmunoassay and the Salimetrics Salivary DHEA EIA. The DHEA serum-saliva correlation, using a log 10 transformation for the total (n = 39), combined males and females is 0.857, $p < 0.0001$.

Method Comparison

The correlation between the Salimetrics EIA and a published serum RIA modified for use with saliva as evaluated by assaying 40 common samples. The EIA-RIA results were highly correlated, $r (38) = 0.881$, $p < 0.001$.

Sample Dilution Recovery

Two samples were serially diluted with Assay Diluent and assayed.

Saliva Sample	Dilution Factor	Expected (pg/mL)	Observed (pg/mL)	Recovery (%)
1			334.66	
	1:2	167.33	157.33	94.0
	1:4	83.67	91.22	109.0
	1:8	41.83	49.44	118.2
	1:16	20.92	20.08	96.0
2			511.38	
	1:2	255.69	287.69	112.5
	1:4	127.84	140.12	109.6
	1:8	63.92	69.30	108.4
	1:16	31.96	34.25	107.2



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com

Antibody Specificity

Compound	Spiked Concentration (ng/mL)	% Cross-reactivity in Salivary DHEA EIA
DHEA-S	1000	0.063
Androstenedione	1000	0.0378
17 β -Estradiol	1	ND
Estriol	1000	ND
Estrone	1000	ND
Progesterone	1000	ND
17 α -Hydroxyprogesterone	1000	ND
Testosterone	1000	ND
Dihydroxytestosterone	1000	ND
Dianabol	1000	ND
11-Hydroxytestosterone	1000	ND
19-Nortestosterone	1000	ND
Cortisol	1000	ND
Aldosterone	1000	ND
Cortisone	1000	ND
11-Deoxycortisol	1000	ND
21-Deoxycortisol	1000	ND
Triamcinolone	1000	ND
Corticosterone	1000	ND
Transferrin	1000	ND

ND = None detected (<0.004)



101 Innovation Boulevard • Suite 302 • State College, PA 16803
 1.800.790.2258 • support@salimetrics.com • salimetrics.com

References

1. Vining, R.F. & McGinley, R.A. (1987). The measurement of hormones in saliva: Possibilities and pitfalls. *J Steroid Biochem*, 27(1-3), 81-94.
2. Reisch, N., Slawik, M., Zwermann, O., Beuschlein, F., & Reincke, M. (2005). Genetic influence of an ACTH receptor promoter polymorphism on adrenal androgen secretion. *Eur J Endocrinol*, 153(5), 711-15.
3. Rainey, W.E. & Nakamura, Y. (2008). Regulation of the adrenal androgen biosynthesis. *J Steroid Biochem Mol Biol*, 108(3-5), 281-86.
4. Sirianni, R., Carr, B.R., Andò, S., & Rainey, W.E. (2003). Inhibition of Src tyrosine kinase stimulates adrenal androgen production. *J Mol Endocrinol*, 30(3), 287-99.
5. Ahn, R.S., Lee, Y.J., Choi, J.Y., Kwon, H.B., & Chun, S.I. (2007). Salivary cortisol and DHEA levels in the Korean population: Age-related differences, diurnal rhythm, and correlations with serum levels. *Yonsei Med J*, 48(3), 379-88.
6. Liu, Y. & Papadopoulos, V. (2005). Biosynthesis and regulation of dehydroepiandrosterone (DHEA) in brain. *Trends de la SCB*, 56, 135-45.
7. Buford, T.W. & Willoughby, D.S. (2008). Impact of DHEA(S) and cortisol on immune function in aging: A brief review. *Appl Physiol Nutr Metab*, 33(3), 429-33.
8. Baulieu, E.E. & Robel, P. (1998). Dehydroepiandrosterone (DHEA) and dehydroepiandrosterone sulfate (DHEAS) as neuroactive neurosteroids. *Proc Natl Acad Sci U S A*, 95(8), 4089-91.
9. Charalampopoulos, I., Alexaki, V.I., Tsatsanis, C., Minas, V., Dermitzaki, E., Lasaridis, I., Vardouli, L., et al. (2006). Neurosteroids as endogenous inhibitors of neuronal cell apoptosis in aging. *Ann N Y Acad Sci*, 1088, 139-52.
10. Pang, S. & Riddick, L. (1990). Hirsutism. In: F. Lifshitz (Ed.), *Pediatric Endocrinology, A Clinical Guide* (2nd ed.), pp. 259-91. New York: Marcel Dekker.
11. Lee, P.D., Winter, R.I., & Green, O.C. (1985). Virilizing adrenocortical tumors in childhood: Eight cases and a review of the literature. *Pediatrics*, 76, 437-44.
12. Morales, A.J., Nolan, J.J., Nelson, J.C., & Yen, S.S. (1994). Effects of replacement dose of dehydroepiandrosterone in men and women of advancing age. *J Clin Endocrinol Metab*, 78(6), 160-67.
13. Manninger, N., Wolkowitz, O.M., Reus, V.I., Epel, E.S., & Mellon, S.H. (2009). Neurobiological and neuropsychiatric effects of dehydroepiandrosterone (DHEA) and DHEA sulfate (DHEA-S). *Front Neuroendocrinol*, 30(1), 65-91.
14. Erb, J.L., Kadane, J.B., Tourney, G., Mickelsen, R., Trader, D., Szabo, R., & Davis, V. (1981). Discrimination between schizophrenic and control subjects by means of plasma dehydroepiandrosterone measurements. *J Clin Endocrinol Metab*, 52, 181-86.
15. Taelman, P., Kaufman, J.M., Janssens, X., & Vermeulen, A. (1989). Persistence of increased bone resorption and possible role of dehydroepiandrosterone as a bone metabolism determinant in osteoporotic women in late post-menopause. *Maturitas*, 11(1), 65-73.
16. Labrie, F., Diamond, P., Cusan, L., Gmez, J.L., Belanger, A., & Candas, B. (1997). Effect of 12-month dehydroepiandrosterone replacement therapy on bone, vagina, and endometrium in postmenopausal women. *J Clin Endocrinol Metab*, 82(10), 3498-3505.
17. De Pergola, G., Giagulli, V.A., Garruti, G., Cospite, M.R., Giorgino, F., Cignarelli, M., & Giorgino, R. (1991). Low dehydroepiandrosterone circulating levels in premenopausal obese women with very high body mass index. *Metabolism*, 40, 187-90.



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com

18. Straub, R.H. & Cutolo, M. (2001). Involvement of the hypothalamic-pituitary-adrenal/gonadal axis and the peripheral nervous system in rheumatoid arthritis: Viewpoint based on a systemic pathogenetic role. *Arthritis Rheumatism*, *44*(3), 493-507.
19. Hucklebridge, F., Hussain, T., Evans, P., & Clow, A. (2005). The diurnal patterns of the adrenal steroids cortisol and dehydroepiandrosterone (DHEA) in relation to wakening. *Psychoneuroendocrinology*, *30*(1), 51-7.
20. Chard, T. (1990). *An introduction to radioimmunoassay and related techniques* (4th ed.). Amsterdam: Elsevier.
21. Schwartz, E.B., Granger, D.A., Susman, E.J., Gunnar, M.R., & Laird, B. (1998). Assessing salivary cortisol in studies of child development. *Child Dev*, *69*(6), 1503-13.
22. Kivlighan, K.T., Granger, D.A., Schwartz, E.B., Nelson, V., & Curran, M. (2004). Quantifying blood leakage into the oral mucosa and its effects on the measurement of cortisol, dehydroepiandrosterone, and testosterone in saliva. *Horm Behav*, *46*(1), 39-46.
23. Schwartz, E., & Granger, D.A. (2004). Transferrin enzyme immunoassay for quantitative monitoring of blood contamination in saliva. *Clin Chem*, *50*(3), 654-56.



Seller's Limited Warranty

"Seller warrants that all goods sold hereunder will be free from defects in material and workmanship. Upon prompt notice by Buyer of any claimed defect, which notice must be sent within thirty (30) days from date such defect is first discovered and within three months from the date of shipment, Seller shall, at its option, either repair or replace the product that is proved to Seller's satisfaction to be defective. All claims should be submitted in written form. This warranty does not cover any damage due to accident, misuse, negligence, or abnormal use. Liability, in all cases, will be limited to the purchased cost of the kit.

It is expressly agreed that this limited warranty shall be in lieu of all warranties of fitness and in lieu of the warranty of merchantability. Seller shall not be liable for any incidental or consequential damages that arise out of the installation, use or operation of Seller's product or out of the breach of any express or implied warranties."

Salimetrics, LLC
5962 La Place Court, Suite 275
Carlsbad, CA 92008, USA
(T) 760.444.4126
(F) 814.238.1608
800-790-2258 (USA & Canada only)
www.salimetrics.com
support@salimetrics.com

Salimetrics, LLC
101 Innovation Blvd., Suite 302
State College, PA 16803, USA
(T) 814.234.2617
(F) 814.238.1608
800-790-2258 (USA & Canada only)
www.salimetrics.com
support@salimetrics.com

"European Authorized
Representative"
Stratech Scientific Ltd
Cambridge House, St Thomas Place,
Cambridgeshire Business Park,
Ely, CB7 4EX, UK
(T) +44 (0) 1638782600
www.stratech.co.uk
info@stratech.co.uk



Updated: March 2, 2018



101 Innovation Boulevard • Suite 302 • State College, PA 16803
1.800.790.2258 • support@salimetrics.com • salimetrics.com